

检测报告

报验编号 232100027230

样品名称 双极水巾

申请人 火人京创(北京)医疗器材有限公司



宁波海关技术中心
Ningbo Customs District Technology Center

声 明

- 1、本单位保证检测的公正性、独立性和诚实性，对检测结果负责，对委托方所提供的检测样品及在检测活动中获得的国家秘密、商业秘密、技术秘密保密。
- 2、本报告无授权签字人签字及未盖本单位检验检测专用章无效。
- 3、本报告涂改、缺页、部分复印无效，复制报告未重新加盖本单位检验检测专用章无效。
- 4、委托方若对本报告有异议，须在检测报告收到之日起十五日内向本单位提出。政府行政管理部门下达的指令性任务，被检方对抽样结果有异议时，应按照政府行政管理部门文件规定及国家相关法律、法规规定进行。
- 5、本单位接收的委托送检样品，其代表性、真实性和准确性由委托方负责。本报告的检测数据和结果仅对收到的样品负责。

宁波海关技术中心

Ningbo Customs District Technology Center

232100027230
2021年12月27日
第1页共1页

检测报告

申请人：火人京创(北京)医疗器材有限公司

申请人地址：北京市大兴区中关村科技园大兴生物医药产业基地天华街二号

生产企业：天津兰天碧水科技有限公司

生产企业地址：天津市武清区城关镇廊良路(城关段)1号

样品名称：双极水巾

规格：24片/盒

批号：20210130

颜色物态：无色透明液体附着白色无纺布上

样品数量：6盒

以上信息及样品由申请人提供及确认，检测单位不承担证实申请人提供信息的准确性、适当性和（或）完整性的责任。

收样日期：2021-12-17

检测日期：2021-12-17~2021-12-27

检测结果：

检测项目	检测结果	检测方法
可迁移性荧光增白剂	无可迁移性荧光增白剂	GB/T 27728-2011 附录 D
pH(25℃)	6.5	GB/T 1545-2008

备注：1.以上检测结果仅对收到的样品负责。未经检测单位同意，申请人不得擅自使用检测结果进行不当宣传。

2.本报告仅用于科研、教学、内部质量控制活动。

授权签字人

检验检测专用章
(16)

地址：宁波市清逸路66号 电话：0574-87022622 投诉：0574-87022702 传真：0574-87022902 邮编：315048 网址：www.nbciqtc.com

声明

- ① 本报告无本单位检验检测专用章无效；
The report is invalid without the special seal for testing.
- ② 本报告无授权签字人签名无效；
The report is invalid unless it is signed by authorized signatory.
- ③ 本报告涂改、缺页、部分复印无效；
The report will be invalid if altered, deficient or partly duplicated.
- ④ 若非本单位抽样，本报告仅对收到的样品负责。
The report is responsible for the received samples only, unless the samples are taken by ourselves.